ECG Equipment Requirements (Updated 11/30/2016)

The FAA does not require a specific type of machine, however the ECG machine used must give a clear picture AND meet the following technical requirements:

- Must generate an image that can be converted to a PDF;

- Must be recorded at 25mm/sec (this is standard in the US);

- Recordings at 50mm/sec will NOT be accepted. Many international programs are set at 50mm/sec as a baseline; the examiner must change this to 25mm/sec for the FAA to accept the tracing; and

- 300 dpi color resolution (or better)